

This information is provided free of charge by the Department of Industrial Relations from its web site at [www.dir.ca.gov](http://www.dir.ca.gov). These regulations are for the convenience of the user and no representation or warranty is made that the information is current or accurate. See full disclaimer at [https://www.dir.ca.gov/od\\_pub/disclaimer.html](https://www.dir.ca.gov/od_pub/disclaimer.html).

Subchapter 7. General Industry Safety Orders  
Group 16. Control of Hazardous Substances  
Article 110. Regulated Carcinogens

[Return to index](#)  
[New query](#)

---

## Appendix A to Section 5204 - Methods of Sample Analysis (Mandatory)

---

This appendix specifies the procedures for analyzing air samples for respirable crystalline silica, as well as the quality control procedures that employers must ensure that laboratories use when performing an analysis required under Section 5204 (d)(5). Employers must ensure that such a laboratory:

1. Evaluates all samples using the procedures specified in one of the following analytical methods: OSHA ID-142; NMAM 7500; NMAM 7602; NMAM 7603; MSHA P-2; or MSHA P-7;
2. Is accredited to ANS/ISO/IEC Standard 17025:2005 with respect to crystalline silica analyses by a body that is compliant with ISO/IEC Standard 17011:2004 for implementation of quality assessment programs;
3. Uses the most current National Institute of Standards and Technology (NIST) or NIST traceable standards for instrument calibration or instrument calibration verification;
4. Implements an internal quality control (QC) program that evaluates analytical uncertainty and provides employers with estimates of sampling and analytical error;
5. Characterizes the sample material by identifying polymorphs of respirable crystalline silica present, identifies the presence of any interfering compounds that might affect the analysis, and makes any corrections necessary in order to obtain accurate sample analysis; and
6. Analyzes quantitatively for crystalline silica only after confirming that the sample matrix is free of uncorrectable analytical interferences, corrects for analytical interferences, and uses a method that meets the following performance specifications:
  - 6.1 Each day that samples are analyzed, performs instrument calibration checks with standards that bracket the sample concentrations;
  - 6.2 Uses five or more calibration standard levels to prepare calibration curves and ensures that standards are distributed through the calibration range in a manner that accurately reflects the underlying calibration curve; and
  - 6.3 Optimizes methods and instruments to obtain a quantitative limit of detection that represents a value no higher than 25 percent of the PEL based on sample air volume.

Note: Authority cited: Sections 142.3, 9020, 9030 and 9040, Labor Code. Reference: Sections 142.3, 9004(d), 9009, 9020, 9031 and 9040, Labor Code.

### HISTORY

1. New Appendix A filed 10-17-2016; operative 10-17-2016 pursuant to Labor Code section 142.3. Submitted to OAL for filing and printing only pursuant to Labor Code section 142.3(a)(3) (Register 2016, No. 43).

 [Go Back to Article 110 Table of Contents](#)